



Effects of Pulsed Electromagnetic Field Therapy at Different Frequencies and Durations on Rotator Cuff Tendon-to-Bone Healing in a Rat Model

Julianne Huegel, PhD¹
Daniel Choi, BS¹
Courtney Nuss¹
Molly Minnig¹
Jennica Tucker, BS¹
Cody Hillin, MD¹
Andrew Kuntz, MD¹
Erik Waldorff, PhD²
Nianli Zhang²
James Ryaby²
Louis Soslowsky, PhD¹

¹McKay Orthopaedic Research Laboratory, Philadelphia, PA

²Orthofix Inc., Lewisville, TX

Introduction

Rotator cuff tears affect millions of individuals each year, often requiring surgical intervention. Although advancements in surgical methods and rehabilitation protocols have improved clinical results, rotator cuff repair failure is common¹. To further improve surgical outcomes, various non-invasive therapeutics have been utilized post-operatively^{2,3}. We have previously shown that pulsed electromagnetic field (PEMF) therapy improved tendon-to-bone healing in terms of tendon modulus in a rat rotator cuff model⁴. While it is known that several cell and tissue responses, including osteogenic differentiation of stem cells, are frequency dependent, the effect of frequency has not yet been evaluated in this system^{5,6}. Therefore, the objective of this study was to determine the influence of both PEMF frequency and exposure time on rotator cuff healing. We hypothesized that a PEMF signal with a higher fundamental frequency and for a longer duration would lead to further improvements in mechanical properties.

Materials and Methods

210 (including 60 from⁴) adult male Sprague-Dawley rats (400-450 g) were used in an IACUC approved protocol. Animals underwent acute supraspinatus injury and repair⁷ followed by either Physio-Stim® PEMF (PS, Orthofix, Inc.) or High Frequency PEMF (HF, similar to PS but with a higher fundamental frequency) for 1, 3, or 6 hours daily. Control animals did not receive PEMF therapy (non-PEMF). Animals were sacrificed at 4, 8, or 16 weeks (n = 10 per group per time point). At sacrifice, right shoulders (n = 7 per group per time point) were dissected and processed for histological analysis, including quantification of fiber alignment circular standard deviation as a measure of collagen organization⁸⁻¹⁰. Left limbs (n = 10 per group per time point) were frozen at -20°C and thawed for dissection prior to tendon cross-sectional area measures and mechanical testing^{7,10,11}. Statistical comparisons were made between control animals and all treatment groups at each time point. Mechanical testing and collagen fiber organization comparisons were made using one-

way ANOVAs with post-hoc tests. Histological comparisons were made using Mann-Whitney U tests. Significance was set at p<0.05.

Results

Mechanical properties

Improvements in mechanical properties were identified for all treatment modalities when compared to non-treated animals (Figure 1A, B). Specifically, one hour of PS treatment led to increased tendon stiffness at all time points, as well as an increase in modulus at 4 weeks. One hour of HF treatment increased stiffness and modulus at 8 weeks. Animals treated with three hours of PS had a decreased cross sectional area (not shown), increased max stress (not shown), and increased stiffness at 4 weeks, as well as increased modulus at 4 and 8 weeks. Three hours of HF PEMF led to increased max load at 4 weeks (not shown), and increased stiffness at 4 and 8 weeks. Treatment with six hours of PS increased modulus substantially at 4 weeks, and increased stiffness and modulus at 8 weeks. Cross-sectional area was also reduced at 4 weeks, but increased at 16 weeks (not shown). The final treatment regimen, 6 hours of HF PEMF, resulted in significant mechanical improvements; including increased stiffness at all time points and a three-fold increase in tendon modulus 8 weeks after repair. Tendon cross sectional area was increased at 16 weeks (not shown). No differences were noted in percent relaxation in any group.

Histological observations

No differences were measured in tendon cell shape or cell density in any treatment group when compared to controls. Collagen organization showed improvements at the tendon insertion at 16 weeks in animals treated with 3 hours of HF PEMF (Figure 2A). This group also showed improved alignment at 8 weeks in the tendon midsubstance (Figure 2B). Additional improvements were identified for 1 hour HF PEMF and 6 hours of PS at 16 weeks in the midsubstance (Figure 2B). Importantly, no adverse effects were identified in any mechanical or histological property.




Figure 1. Tendon mechanical properties. **(A)** Tendon stiffness was increased in all groups receiving PEMF treatment, regardless of modality, when compared to non-PEMF controls. **(B)** Modulus was improved in all groups except HF (3 hr) at 4 and/or 8 weeks. By 16 weeks, properties were similar between control and treatment groups. Data displayed as mean \pm SD. Bars indicate $p < 0.05$.

Figure 2. Collagen fiber alignment. **(A)** At 16 weeks, collagen alignment was improved at the injury site (insertion) in animals treated with HF (3 hr). **(B)** Collagen alignment was similarly improved for HF (3 hr) at 8 weeks in the midsubstance. Improvements were also seen at 16 weeks in groups treated with HF (1 hr) and PS (6 hr). Data displayed as mean \pm SD. Bars indicate $p < 0.05$.

Discussion

Overall, results suggest that PEMF has a positive effect on rat rotator cuff healing for pulse frequency or treatment duration tested in this study. Tendon mechanical properties, including tendon stiffness and modulus, were significantly improved with all six treatment modalities compared with control. Additionally, collagen fiber organization was improved after treatments, suggesting more organized tissue in PEMF-treated tissues, consistent with the improved mechanical strength seen in treated groups. We suspect that PEMF treatment may increase tendon cell metabolism, which then in turn increases both collagen production and matrix remodeling. These proposed changes are supported by our findings of improved mechanical properties and improved collagen alignment. Our previous work indicates that PEMF treatment does not alter joint function⁴; in conjunction with these current findings, these animal studies promote the evaluation of PEMF to improve rotator cuff healing in the clinical setting.

Conclusions

Non-invasive PEMF therapy improves tendon-to-bone healing in an acute rat supraspinatus detachment and repair model, supporting investigation of this treatment in a clinical scenario of post-operative rotator cuff healing.

Acknowledgements

Funding was provided by Orthofix, Inc. The authors thank Stephanie Weiss and Jessica Johnston for assistance with surgical procedures, and Harina Raja for assistance with histological analysis.

References

1. Galatz LM, Ball CM, Teefey SA, Middleton WD, Yamaguchi K. The outcome and repair integrity of completely arthroscopically repaired large and massive rotator cuff tears. *J Bone Joint Surg Am.* 2004 Feb;86-A(2):219-24. PubMed PMID: 14960664.

- Lovric V, Ledger M, Goldberg J, Harper W, Bertollo N, Pelletier MH, Oliver RA, Yu Y, Walsh WR. The effects of low-intensity pulsed ultrasound on tendon-bone healing in a transosseous-equivalent sheep rotator cuff model. *Knee Surg Sports Traumatol Arthrosc.* 2013 Feb;21(2):466-75. doi: 10.1007/s00167-012-1972-z. Epub 2012 Mar 31. PubMed PMID: 22466014.
- Springer J, Badgett RG. ACP Journal Club: optimized extracorporeal shock-wave therapy improved pain and functioning in chronic plantar fasciitis. *Ann Intern Med.* 2015 Nov 17;163(10):JC8. doi: 10.7326/ACPJC-2015-163-10-008. PubMed PMID: 26571262.
- Tucker JJ, Riggan CN, Connizzo BK, Mauck RL, Steinberg DR, Kuntz AF, Soslowky LJ, Bernstein J. Effect of overuse-induced tendinopathy on tendon healing in a rat supraspinatus repair model. *J Orthop Res.* 2016 Jan;34(1):161-6. doi: 10.1002/jor.22993. Epub 2015 Aug 19. PubMed PMID: 26218457; PubMed Central PMCID: PMC4710550.
- Miller SL, Coughlin DG, Waldorff EI, Ryaby JT, Lotz JC. Pulsed electromagnetic field (PEMF) treatment reduces expression of genes associated with disc degeneration in human intervertebral disc cells. *Spine J.* 2016 Jun;16(6):770-6. doi: 10.1016/j.spinee.2016.01.003. Epub 2016 Jan 15. PubMed PMID: 26780754.
- Luo F, Hou T, Zhang Z, Xie Z, Wu X, Xu J. Effects of pulsed electromagnetic field frequencies on the osteogenic differentiation of human mesenchymal stem cells. *Orthopedics.* 2012 Apr;35(4):e526-31. doi: 10.3928/01477447-20120327-11. PubMed PMID: 22495854.
- Beason DP, Connizzo BK, Dourte LM, Mauck RL, Soslowky LJ, Steinberg DR, Bernstein J. Fiber-aligned polymer scaffolds for rotator cuff repair in a rat model. *J Shoulder Elbow Surg.* 2012 Feb;21(2):245-50. doi: 10.1016/j.jse.2011.10.021. Review. Erratum in: *J Shoulder Elbow Surg.* 2013 Apr;22(4):581. PubMed PMID: 22244068.
- Gimbel JA, Van Kleunen JP, Williams GR, Thomopoulos S, Soslowky LJ. Long durations of immobilization in the rat result in enhanced mechanical properties of the healing supraspinatus tendon insertion site. *J Biomech Eng.* 2007 Jun;129(3):400-4. PubMed PMID: 17536907.
- Bey MJ, Song HK, Wehrli FW, Soslowky LJ. A noncontact, nondestructive method for quantifying intratissue deformations and strains. *J Biomech Eng.* 2002 Apr;124(2):253-8. PubMed PMID: 12002136.
- Thomopoulos S, Williams GR, Gimbel JA, Favata M, Soslowky LJ. Variation of biomechanical, structural, and compositional properties along the tendon to bone insertion site. *J Orthop Res.* 2003 May;21(3):413-9. PubMed PMID: 12706013.
- Gimbel JA, Van Kleunen JP, Mehta S, Perry SM, Williams GR, Soslowky LJ. Supraspinatus tendon organizational and mechanical properties in a chronic rotator cuff tear animal model. *J Biomech.* 2004 May;37(5):739-49. PubMed PMID: 15047003.